



University of Chemistry and Technology, Prague
Metrological and Testing Laboratory UCT Prague

Testing laboratory No. 1316.2 accredited by the CAI according to the EN ISO/IEC 17025:2018



Address: VSCHT Praha, Technická 1905/5, 166 28 Prague 6, Czech Republic (tel.: +420 602833424; +420 220443184; https://www.vscht.cz/mzl)

Test certificate ML: 975/25

print no.: ENG_409/25

Client: EUPHORIA TRADE s.r.o.

Klimentská 1216/46
110 00 Praha - Praha 1
Česká republika

Sample received: 19.5.2025
Order no.: 15.5.2025
Sample description (client's): T9HC Gum mies

Testing item: Gum mies
packaging: polyethylene bag (PE)
quantity: 1 g

Date of testing: 19.5.2025 - 26.5.2025
Location of testing: facilities of the MZL UTC, Technická 1903/3, 166 28 Prague 6 - Dejvice
Testing methods used: KM 21: LC-MS

TEST RESULTS:

CANNABINOIDS

Analyte	Result*	Expanded uncertainty	Unit	Testing method	Notice
CBD (cannabidiol)	<0.00005	-	% w/w	KM 21	
CBDA (cannabidiolic acid)	<0.00005	-	% w/w	KM 21	
trans- Δ^9 -THC (trans-delta-9-tetrahydrocannabinol)	<0.00005	-	% w/w	KM 21	
Δ^9 -THCA-A (delta-9-tetrahydrocannabinolic acid-A)	<0.00005	-	% w/w	KM 21	
Δ^9 -THC equivalents (sum of Δ^9 -THC + Δ^9 -THCA-A x 0.877)	<0.00005	-	% w/w	KM 21	

* the sign "<" indicates that concentration is lower than this value, i.e. below the limit of quantitation (LOQ)

Expanded uncertainty was calculated using coverage factor $k = 2$ corresponding to a coverage probability of approximately 95%.
Uncertainty was calculated and stated according to the ILAC G17:01(2021) and Kvalimetrie 11 (EURACHEM/CITAC 4). Uncertainty of sampling is not covered.

The results given herein apply only to the sample as received. This certificate shall not be reproduced except in full, without written approval of the Laboratory. The certificate does not substitute any other legal document. Laboratory is not responsible for information supplied by customer, if such information can affect the validity of results.

Appendix:

Date of issue: 26.5.2025

prof. Jana Hajslova, Ph.D., Head of Laboratory

The end of Certificate